



2016 NIST Handbook 44 Section 1.10. General Code G-UR. User Requirements

G-UR.4.2. Abnormal Performance. – Unstable indications or other abnormal equipment performance observed during operation shall be corrected and, if necessary, brought to the attention of competent service personnel. (Added 1976)